



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

PRECISION GRANITE  
353 Plunkett Creek Rd  
Gordonsville, TN 38563  
Darren Smith Phone: 615 604 5673  
prgranite@yahoo.com

CALIBRATION

Valid To: February 29, 2028

Certificate Number: 1679.01

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with R205 – A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Dimensional

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
Surface Plates <sup>3</sup> – Flatness	(8 × 12) in to (10 × 10) ft	5.0 μin/ft	GGGP 463c Autocollimator Note: CMC is to be no less than the acceptable closure error for the procedure
Repeat Reading	0.002 in	20 μin	Repeat-o-meter – only valid in connection with a flatness calibration

<sup>1</sup> This laboratory offers commercial and field calibration service only.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

<sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.



# Accredited Laboratory

A2LA has accredited

## PRECISION GRANITE

*Gordonsville, TN*

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 12<sup>th</sup> day of January 2026.

A blue ink signature of Trace McInturff.

Mr. Trace McInturff Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1679.01  
Valid to February 29, 2028

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*